

A Central Laboratory Inter-laboratory Comparison Program to Assess the Comparability of Data of Forty-one Tests from Four Regional Laboratories Involved in Global Clinical Trials over a Twelve-Month Period

Christopher R. Daniels, PhD; Venkat R. Rajagopal, MS; Traci Turner, MD
Medpace Reference Laboratories, Cincinnati, OH, USA

ABSTRACT

Objective

The objective of this study was to develop and support a process for assessing the comparability of data used in global clinical trials from four individual laboratories, owned by the same central laboratory entity, and that the same tests on the same samples would be under statistical control and acceptable limits of variation.

Relevance

Data from central laboratories have been key for assessing safety, tolerability, and efficacy of new drugs in clinical trials. With the increasing complexity and global scale of many clinical trials, it is important to maintain harmonization among the regional laboratories as part of a central laboratory participating in the same study. However, continuous monitoring of the same samples run at all laboratory locations may not be common practice.

Methodology

Pooled serum, plasma, and urine samples were aliquoted, frozen at -70°C, and distributed quarterly to each laboratory. Samples were analyzed weekly on the same day at each laboratory for twelve months. The percent and absolute bias were calculated for each result using the US laboratory as the reference laboratory. The percent bias for each week and month was also calculated. A Bland-Altman plot was created between each laboratory and the reference laboratory for the twelve-month period, and a student T-test was run using a relative bias limit for each test and a significance level of 5%.

Results

Overall, during the twelve month period, all forty-one tests had a twelve month mean bias within the acceptable bias limit for each individual test compared to the reference laboratory. When comparing the mean weekly bias for each tests from all laboratories, any week where the bias was outside acceptable limits, investigation and corrective action was undertaken to determine the source of the error. Examples include:

- One week during the six-month period, three chemistry tests from the same laboratory had a mean weekly bias outside the acceptable limit. After investigation, it was determined that a pre-analytical issue with thawing/mixing of frozen samples was the cause of the bias.
- A negative bias was observed in a US laboratory for triglyceride. However, subsequent review of CDC Lipid Standardization Program Part III data from all laboratories globally demonstrated acceptable performance.
- Inconsistencies in the reporting of results across laboratories, specifically the technical decision not to report data deemed biologically implausible, were noted after monthly review of calcium data, and all technologists were re-educated on handling repeat analysis and appropriate consultation of Laboratory Directors.


Conclusion

An inter-laboratory program where frequent monitoring of identical samples run at all laboratories involved in clinical trials can provide valuable information into the harmonization of data reported by the central laboratory and help mitigate pre-analytic, analytic, and post-analytic issues that may arise when assessing data used in the development of new therapeutics.

Table 1: List of Analytes and Platforms			
Beckman Coulter AU Series Chemistry Analyzers			
ALP (U/L)	Lipoprotein (a) (mmol/L)		
ALT (U/L)	Iron (µg/dL)		
AST (U/L)	BUN (mg/dL)		
Amylase (U/L)	Creatinine (mg/dL)		
Chloride (mmol/L)	UA (mg/dL)		
Creatinine Kinase (U/L)	PO4 (mg/dL)		
LDH (U/L)	Sodium (mmol/L)		
GGT (U/L)	Triglyceride (mg/dL)		
TBil (mg/dL)	Total Cholesterol (mg/dL)		
Albumin (g/dL)	HDL-Cholesterol (HDL-C) (mg/dL)		
Total Protein (g/dL)	Magnesium (mg/dL)		
Calcium (mg/dL)	Urine Creatinine (mg/dL)		
Glucose (mg/dL)	Urine Protein (mg/dL)		
Potassium (mmol/L)			
Roche Immunoanalyzers		Siemens BNII Nephelometer	
Insulin (µIU/mL)	Apo AI (mg/dL)		
TSH (µIU/mL)	Apo B (mg/dL)		
T4 (µg/dL)	hsCRP (mg/L)		
C-peptide (ng/mL)	Urine Albumin (mg/dL)		
Stago Compact		Tosoh G7/G8	
PT (sec)	HbA1C (%)		
aPTT (sec)			
Preparative Ultracentrifugation (PUC)			
LDL Cholesterol (LDL-C) (mg/dL)			
VLDL Cholesterol (VLDL-C) (mg/dL)			
VLDL-C/Trig ratio			

Statistical methods

- Summary statistics listing the bias criteria, number of samples, and mean difference (absolute or relative) compared to the reference laboratory (MRL-US) were included. The P-values are from TOST t-test performed on absolute or relative difference.
- A Bland-Altman plot was created between each laboratory and reference laboratory for the 12-month period (Figure 1).
- Bias criteria were based on criteria from the College of American Pathologists, Westgard Biological Variation Database, and Royal College of Pathologists of Australasia.
- Some test and laboratory locations have an N of <520 for a given analyte due to insufficient quantity, instrument error, etc.



RESULTS

All tests run on Beckman Coulter AU Series Chemistry Analyzers, Roche Immunoanalyzers, Siemens BNII Nephelometers, Stago Compact Coagulation Analyzers, Tosoh G7/G8 HPLC Analyzers, or by preparative ultracentrifugation demonstrated acceptable equivalence when compared to the reference laboratory over a 12-month period (Table 2). Representative examples of pre-analytic, analytic, and post-analytic issues detected over the twelve-month period include:

- Monthly review of BUN data (Table 3) showed a mean bias for BE laboratory of -14.28%. Upon further review of other tests run on the same sample, a similar negative bias was seen (Figures 2, 3, 4). After investigation, it was determined the cause of bias was due to a pre-analytical error, specifically improper mixing during the freeze/thaw process. Technologists were re-educated on the proper freeze/thaw process when analyzing samples that have been stored at -70°C.
- A negative bias of ~4.5% for triglyceride was observed in the US laboratory compared to the other laboratory locations (Figures 5, 6, 7). The CDC Lipid Standardization program results for 2016 were reviewed for all four laboratories and indicated good performance, with the bias averaging -2.29%, -0.51%, -0.57%, and -1.76% for the US, BE, SG, and CN labs, respectively, when compared to CDC targets (Table 4).
- Reporting inconsistencies were noted between laboratories for calcium during the review period, with some labs reporting a numeric value, and other labs electing to not report a numeric value due to concerns the calcium values were biologically implausible, indicating a pre-analytic error affecting sample integrity (data not shown).

Corrective actions included re-education on the process to determine the validity of results, including, but not limited to, repeat analysis, delta checking, and consultation with the Laboratory Director as to clinical validity.







CONCLUSION

An inter-laboratory comparison program where frequent and continuous monitoring of identical samples analyzed at all laboratories involved in clinical trials is conducted can provide valuable information into the harmonization of data reported by the central laboratory and help mitigate pre-analytic, analytic, and post-analytic issues that may arise when assessing data used in the development of new therapeutics.

Table 2: Summary Statistics of Analytes							
Test	Bias Criteria	BE		CN		SG	
		N	Mean Diff. (Abs or %)	N	Mean Diff. (Abs or %)	N	Mean Diff. (Abs or %)
Beckman Coulter Chemistry Analyzers							
ALP	12%; Abs (15) if ≤125	512	2	513	-3	509	-2
ALT	10%	510	-2.3%	520	-3.9%	516	5.7%
AST	10%	510	-1.0%	520	-1.9%	516	4.2%
Albumin	6%; Abs (0.2) if ≤3.3	510	0.4%	520	2.1%	516	2.3%
Amylase	10%	510	1.8%	520	-0.1%	516	2.2%
BUN	10%	510	0.1%	520	-1.8%	516	2.2%
Calcium	4%	499	0.8%	510	2.5%	508	0.2%
Chloride	2.5%	510	0.5%	520	0.2%	516	0.1%
Creatinine	10%	510	0.9%	520	5.0%	516	4.0%
Creatine Kinase	8%	510	-0.4%	520	0.7%	516	-0.5%
Gamma-Glutamyl Transferase	8%	510	2.4%	519	1.5%	516	-1.0%
Glucose	5%	510	0.8%	520	1.5%	516	1.9%
Iron	6%	498	1.0%	507	-0.1%	507	1.3%
LDH	7.5%	510	-3.0%	520	0.3%	516	-0.7%
Magnesium	10%	502	1.6%	512	-0.2%	508	1.0
PO ₄	8%; Abs (0.2) if ≤2.3	510	0.1	520	0.1	516	0.1
Potassium	3%	498	0.9%	509	1.6%	507	0.2%
Sodium	2%; Abs (3) if ≤150	510	1.2%	520	0.9%	516	0.7%
Total Bilirubin	12%; Abs (0.18) if ≤1.46	520	0.0%	520	-2.9%	516	5.9%
Total Protein	5%	510	5.0%	520	1.3%	516	0.6%
Triglyceride	8%	510	4.2%	520	2.1%	516	3.9%
	6%	175	0.1%	176	0.8%	173	1.7%
Total Cholesterol	Abs (11.5) if ≤193	337	0.6	346	1.3	343	2.7
HDL-Cholesterol	9.24%	510	0.1%	520	2.9%	516	5.5%
Lipoprotein (a)	15%	518	1.9%	517	3.4%	514	5.1%
Uric Acid	6%	510	2.5%	520	0.5%	516	3.1%
Urine Creatinine	14%	520	-0.5%	520	6.5%	520	-1.1%
Urine Protein	20%	518	-5.2%	518	5.3%	518	0.0%
Insulin	12%	520	-0.2%	520	1.0%	516	3.6%
Thyroid Stimulating Hormone	10%	510	-1.2%	519	1.5%	516	2.6%
Roche Immunoanalyzers							
T4	10%; Abs (0.93) if ≤9.32	510	-1.9%	CN laboratory did not perform T4 Analysis		516	-3.6%
C-Peptide	10%	513	0.6%	519	-0.6%	516	-2.3%
Siemens BNII Nephelometers							
Apolipoprotein AI	10%; Abs (20) if ≤200	510	-3	520	10	516	-3
Apolipoprotein B	10%; Abs (20) if ≤200	510	-3	520	4	516	2
High Sensitivity CRP	10%	430	-0.9%	437	4%	435	-4.0%
Urine Albumin	20%	153	-2.7%	154	7%	153	2.3%
	Abs (0.4) if ≤2	324	0.063	154	2.34	335	0.06
Preparative Ultracentrifugation (PUC)							
LDL-C	10%	49	-3.6%	49	-1.2%	SG laboratory did not perform PUC	
VLDL	40%	49	10.0%	49	2.6%		
VLDL/Trig Ratio	40%	49	5.3%	49	-0.9%		
Stago Compact Coagulation Analyzers							
PT	10%	108	4.9%	CN laboratory did not perform PT or PTT analysis		107	3.4%
PTT	10%	109	7.9%			110	-0.4%
Tosoh G7/G8							
HbA1C	6%	120	-0.9%	100	3.3%	120	1.8%
All P-values are from TOST t-test performed on absolute or relative difference: <0.0001							

Table 3: Weekly Bias for BUN									
Week of	% Bias Weekly Avg (BE)			% Bias Weekly Avg (CN)			% Bias Weekly Avg (SG)		
	BUN	K'	CI	BUN	K'	CI	BUN	K'	CI
31-Jan-2016	-2.89	2.28	1.81	-5.31	1.39	0.49	-0.04	0.22	-0.10
07-Feb-2016	-0.76	3.74	-0.46	-4.19	1.35	-2.54	2.84	0.66	-3.02
14-Feb-2016	-14.28	-11.95	-10.61	0.51	1.80	1.14	2.52	0.93	0.10
21-Feb-2016	-2.20	-4.77	-5.01	-0.69	0.92	0.67	5.87	0.87	0.29
28-Feb-2016	-2.88	1.81	0.57	0.05	1.37	0.19	0.54	0.91	-0.29

Table 4: CDC Lipid Standardization Program 2016									
Pool Series	Triglycerides				Cholesterol				
	US		BE		SG		CN		
	mg/dL	% Bias	mg/dL	% Bias	mg/dL	% Bias	mg/dL	% Bias	
Q1 136	471	116.35	-0.89	119.00	1.37	117.75	0.31	116.75	-0.55
	141	105.73	-2.34	109.00	0.68	108.00	-0.24	92.50	1.38
	472	93.68	-2.67	96.25	0.01	95.50	-0.77	79.75	-3.89
Q2 137	144	78.40	-4.76	81.50	-1.00	80.75	-1.91	107.75	-3.12
	471	106.35	-1.76	108.25	-0.01	107.75	-0.47	107.75	-0.47
	473	227.93	2.84	229.75	3.66	226.00	1.97	230.50	4.00
Q3 138	474	78.03	-4.87	79.25	-3.38	79.50	-3.07	76.50	-6.73
	571	87.15	-2.55	87.00	-2.72	87.50	-2.16	85.50	-4.39
	801	112.95	-3.02	116.75	0.24	116.50	0.03	115.00	-1.26
Q4 139	476	122.63	-3.12	126.25	-0.25	126.50	-0.06	123.50	-2.43
	477	262.18	0.01	260.00	-0.82	267.25	1.95	264.50	0.90
	571	85.58	-4.31	86.00	-3.84	87.25	-2.44	87.75	-1.88
Average Bias %		-2.29		-0.51		-0.57		-1.76	



REFERENCES

- Clinical and Laboratory Standards Institute (CLSI) Estimation of Total Analytical Error for Clinical Laboratory Methods; Approved Guideline EP-21A Volume 23 Number 20
- Westgard Biological Variation Database specifications, updated in 2014. <https://www.westgard.com/biodatabase1.htm>
- National Cholesterol Education program (NCEP), Recommendations on Lipoprotein measurement by the Working Group on Lipoprotein Measurement. (Sept 1995) NIH pub:95-3044
- National Glycohemoglobin Standardization Program (NGSP) (2009). CAP Survey limits lowered 2011.
- Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. Lancet 1986; 1: 307-310.
- <https://www.rcpaq.com.au/wp-content/uploads/2013/06/chempath/Allowable%20Limits%20of%20Performance.pdf>

ACKNOWLEDGMENTS

The authors gratefully acknowledge Megan Coffman for her assistance in the design and layout of this presentation.