



Certificate of Traceability

Level I Laboratory Certification

This certifies that *Medpace Reference Laboratories*, using *Tosoh G8* has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: G8	Calibrator Lot: ZS8002	Column Lot: EA-0044N
Reagent Lot: H8-153E, H8-253E, H8-353E, HW-600-NU	Calibrator Assigned Values: 5.80%, 10.70%	

Date of Certification: July 1, 2020

Certification Expires: July 1, 2021

Daniel Sacks

NGSP Steering Committee Chair

Randie R Little PhD.

NGSP Network Coordinator

Sharon Anulf

SRL director/ supervisor